

03.05.2024

Urgent Field Safety Notice

Dear Customers,

DH Healthcare GmbH, a Dedalus Group company, would like to bring to your attention the following issue reported to the national competent authority:

Title: Consecutive (or Alternate) mode abnormally available when prescribing a prescription set

Internal Reference: MST0079721

Product name and version(s) and UDI-DI:

- ORBIS Medication 03.16.07.00 in ORBIS 84.38.00.09.FR and 85.21.00.10.FR and higher in France.
 Manufacturer: DH Healthcare GmbH UDI-DI: 4260693990026
- ORBIS Medication 03.19.01.00 in ORBIS 84.41.00.00.DACHL and higher in Germany, Austria, Switzerland, Luxembourg.
 Manufacturer: DH Healthcare GmbH UDI-DI: 4260693990026

Information:

The following behavior, which might occur when using ORBIS Medication, has been observed: A physician adds a prescription set to a patient case, with a prescription line to be administered as a daily dose, with no duration.

Before signing, he modifies the prescription line and selects the Consecutive or Alternate mode. He completes the dosage for each step.

URGENT FIELD SAFETY NOTICE - MST0079721 DH Healthcare GmbH Konrad-Zuse-Platz 1-3, 53227 Bonn

 1/5



HOSPITA	L					Measured weight 🖌 1.	7 kg so 0.1505 m² 🥖	P
Prescri	ption [Indications						
oduct(s) &	& Dose	_						
μ		halationslösung, 5 ösung für einen	1 Amp/take					
		Route:	Respiratory (Inhalation)					
		Propellant:	Luft					
							• • • • • • • • • • • • • • • • • • •	
ditional i	information						Comment to the r	nun
		Instructions:	Modify					
nedule								
		Repetition (next days):	Consecutive					
		Reas						
	e prescription							
•	Step 1	Start - End		Duration 1d	Dose	Daily repetition	Condition	
102		Tue Apr 09 1:55 PM			1 Amp/take 2 Amp/take	3 times / day 3 times / day	Modify Modify	
		Wed Apr 10 1:55 Pl						
ίQ.	2	Wed Apr 10 1:55 PI	n - mu Aprill 1.54 PM	1d		5 times / day		
Add ste		Wed Apr 10 1:55 P!		10		5 times / day		
Add ste		Wed Apr 10 1:55 P!	n - Thu Api TT 134 FM	10		s and s add		
Add ste		Wed Apr 10 1:55 Pf	n - nu april 1347m		© 13:55	S times y day		
Add ste		Wed Apr 10 1:55 Pf	1 Amp at 07:00	1 Amp at 12:00	€ 13:55 € 13:55	Amp at 18:00		
Add ste	q.		1 Amp at 07:00	1 Amp at 12:00	• 13:55 13:55 14	Amp at 18:00	0000	
Add ste				1 Amp at 12:00	€ 13:55 € 13:55	Amp at 18:00	00:00 Night	

The prescription line is then only replaced by the last step defined with the Consecutive or Alternate mode The other steps are not considered.

New	lines - To be	e signed				1
•	B	Inhalationslösung, 5 2 - 2 - 2 Amp mg/1 ml Lösung für einen Vernebler, 10 ml aprizh_fhir Aerosol	Propellant: Luft	Salbutamol R03AC02	► 10/04/2024 18:00 ■11/04/2024 12:00	1 d
		Respiratory (Inhalation)				

This could lead to wrong or missing prescriptions for the patient.

URGENT FIELD SAFETY NOTICE - MST0079721

DH Healthcare GmbH Konrad-Zuse-Platz 1-3, 53227 Bonn



Actions:

Actions undertaken by DH Healthcare GmbH:

- Inform the affected customers with this letter.
- Release of correction with ORBIS Medication version 03.20.00.00 in ORBIS version 84.42.00.00 (release beginning of June 2024 for DACHL and planned for spring 2025 for FR) and higher, 85.25.00.00 (release planned for second half of 2025) and higher.

Recommended actions to be taken by the customer:

- As a physician, always check before signature the result of the modification of a prescription line.
- When using an affected version, if consecutive (or alternating) prescriptions are to be prescribed, add a new prescription line directly, by selecting a drug, instead of selecting a protocol.
- Installation of the correction when available.

Please distribute this information to all those who need to be aware of it and confirmation the acknowledgement by the signed response form.

Regardless of the situation described here, we would like to point out that care providers must always ensure that clinically relevant information, including prescription information, is clearly communicated and that they must use verified information (e.g., from medical devices such as monitoring systems), independent from the software being used.

> URGENT FIELD SAFETY NOTICE - MST0079721 DH Healthcare GmbH Konrad-Zuse-Platz 1-3, 53227 Bonn



It is important that you take the actions described in this safety information and acknowledge receipt of this letter.

If the above information does not apply to your hospital or if the device has been transferred to another organization, please indicate this on the attached feedback form and forward this Field Safety Notice to the respective organization.

Thank you for your careful attention to this matter and for your support.

If you have any questions on this matter, please consult our contact person:

Sincerely,

URGENT FIELD SAFETY NOTICE - MST0079721 DH Healthcare GmbH Konrad-Zuse-Platz 1-3, 53227 Bonn



Urgent Field Safety Notice

Feedback Form

We kindly ask you to return this feedback form as soon as possible, but at the latest **within 30 days** after receipt of this letter, to the following e-mail address: Thank you for your cooperation.

Customer / Facility (names of all affected operational facilities):

□ I confirm that I have received and understood the safety information.

□ The safety information does not apply to my facility.

□ The device was transferred to another organization.

Name and address of the other organization: _____

□ Please update our contact information as follows:

Customer / Facility:

Address:

URGENT FIELD SAFETY NOTICE - MST0079721

DH Healthcare GmbH Konrad-Zuse-Platz 1-3, 53227 Bonn

 $//// \lambda$ $| | | //// <math>\sim$ LIFE FLOWS THROUGH OUR SOFTWARE