

URGENT Field Safety Notice

PIC iX Customized Event Catalog Setting Not Copied as Expected From One Unit to Another

<Date of letter deployment,> *<date format: DD-MMM-YYYY, e.g. 02-JAN-2021>*

*<To: Name / Title / Customer Name
Street Address
City, State, Zip Code
<modify title block format as needed>*

This document contains important information for the continued safe and proper use of your equipment.

Please review the following information with all members of your staff who need to be aware of the contents of this communication. It is important to understand the implications of this communication.

Please retain this letter for your records.

Dear Customer,

Philips has become aware of a potential safety issue related to Patient Information Center (PIC) iX customized Event Catalog configuration not copied as expected from one unit to another. This notification is intended to inform you about:

1. The problem and under what circumstances it can occur

PIC iX Event Notification provides healthcare professionals with supplemental information related to alarms and system information messages to suit clinical workflows.

PIC iX Event Catalog tool allows the user to configure categories/events per Unit and define and manage the list of events associated with vendor products. This means that events and event categories associated with a product can be added, edited or deleted to customize the event notifications sent to caregiver mobile devices.

The user can also configure Event Categories and Subcategories to be sent to specific roles at specific escalation levels and customize Event Categories in the Event Catalog in one unit and copy those customizations to other units.

It has been discovered that when the user copies the customized Event Categories from one unit to another, the Event Catalog displays the customized Event Categories, but the changes are not actually applied. If a user reviews the events in the Event Monitor the user will see in the event details that the Event Category is not updated to reflect the customization.

This issue is the result of cached changes to the Event Catalog not propagating in the database.

2. Hazard/harm associated with the issue

If event catalog setting is not copied as expected, there is the potential that clinical users are not properly notified of changes in patient condition. This may lead to the delay in treatment. Although unlikely, this could potentially result in patient harm.

3. Affected products and how to identify them


PIC iX with Event Notification
Software versions 4.x

#	Product Name	Product Code(s)
1	PIC iX	866389
2	PIC iX Expand	866390

Use the following instructions to identify the software revision of PIC iX:

Access the **Product Support** screen by clicking the **Philips** icon in your PIC iX application. The PIC iX Software Serial Number and Software Version appear on the **Product Support** screen in the **Product Information** Section under **Serial Number** and **Software Version** respectively. Refer to Image 1 below:

Image 1: Software Version Identification



4. Actions that should be taken by the customer / user to prevent risks for patients or users

- Review events in Event Monitor and confirm the expected Category is assigned for the event.
- If events are not assigned to the correct category do not depend on event notification for patient monitoring.
- This notice should be passed on to all those who need to be aware within your organization or to any organization where affected devices have been transferred.

5. Actions planned by Philips to correct the problem

A Philips representative will contact you to schedule a visit from a Philips Field Service Engineer who will provide software patch PIC iX 4.2.4 to correct this issue.

If you need any further information, please contact your local Philips representative: *<Philips representative contact details to be completed by the Market/Business>*



Philips regrets any inconvenience caused by this problem.

Sincerely,

Deborah Currlin,
Head of Quality

URGENT Field Safety Notice Response Form

Reference: CR # 2024-CC-HPM-013, PIC iX Customized Event Catalog Setting Not Copied as Expected From One Unit to Another

Instructions: Please complete and return this form to Philips Healthcare promptly upon receipt and no later than 30 days from receipt by email: recall.response@philips.com. Completing this form confirms receipt of the URGENT Field Safety Notice, understanding of the issue, and required actions to be taken.

Customer/Consignee/Facility Name: _____

Street Address: _____

City/State/ZIP/Country: _____

Customer Actions:

- Review events in Event Monitor and confirm the expected Category is assigned for the event.
- If events are not assigned to the correct category do not depend on event notification for patient monitoring.
- Pass this notice to all those who need to be aware within your organization or to any organization where affected devices have been potentially transferred.

We acknowledge receipt and understanding of the accompanying Product Notice and confirm that the information from this Notification has been properly distributed to all users that handle affected devices.

Name of person completing this form:

Signature: _____

Printed Name: _____

Title: _____

Telephone Number: _____

Email Address: _____

Date (DD / MMM / YYYY): _____