

Customer

**Ralf Quentin**  
Manager Quality Management  
and Product Development

Tel.: +49 241 / 91 30 0  
Fax.: +49 241 / 91 30 106  
Email: vigilance@vygon.com

Aachen, 20.08.2024

**Urgent Recall on multicath2, Code 1202.062, Batch No. 170624GR**

We have identified a batch-related manufacturing problem, which may lead to a leakage on the catheter tube.

The following product is involved in this Recall 007:

Product Description	Code No.	Batch No.
multicath2	1202.062	170624GR

We therefore ask you to immediately withdraw this batch from services and to return the batch according to our instructions below.

Please acknowledge receipt of this letter and complete and return the attached form by indicating the quantities withdrawn from your institution.

Your National Competent Authority (ANSM) has been informed about this Recall 007.

If you should require further information, please contact your local Vygon distributor.

We apologize for any inconvenience this recall may cause.

Yours sincerely,

Ralf Quentin  
Manager Quality Management and Product Development

---

**Please complete and return this form:**  
by fax: +49 241 9130 106 or e-mail to : [vigilance@vygon.com](mailto:vigilance@vygon.com)

---

<b>Name and address of the institution :</b>	
<b>Full name of the person to contact:</b>	
<b>Function :</b>	
<b>☎ Phone number:</b>	
<b>✉ E-mail:</b>	

**We acknowledge receipt of the above Recall and that the information contained in this field safety corrective action has been shared with all recipients / end users of above-mentioned products within our organization.**

**Please tick the appropriate box:**

**We have the following product available on stock :**

**Code 1202.062, Batch 170624GR**  Yes  No

**If you have stock, number of units removed: \_\_\_\_\_ Please return these products back to VYGON.**

**Signature and Date :** \_\_\_\_\_