



Urgent Field Safety Notice (Removal)

Cordis SELUTION SLR™ 018 PTA 3.0mm X 100mm Balloon Catheters
For specific lots - See listing in Table below:

Item/Product Number	Lot Number
SE18030100	L92505
SE18030100	L92835
SE18030100	L93029
SE18030100	L93495

October 29, 2024

Dear Valued Customer,

The purpose of this communication is to inform you that, Cordis is voluntarily removing specific lots of SELUTION SLR™ 018 PTA 3.0mm X 100mm Balloon Catheters. You are receiving this letter because our records indicate that you have purchased and have in your possession one or more of the impacted lots of the subject product: SELUTION SLR™ 018 PTA 3.0mm X 100mm Balloon Catheters.

Field Safety Notice (Removal) Overview:	<p>Cordis has identified a potential for "slow deflation" to occur during use of certain lots of SELUTION SLR™ 018 PTA 3.0mm X 100mm. A total of three complaints have been received by Cordis.</p> <p>The potential impacts include but are not limited to situations of patient discomfort, increased procedure time, additional intervention, vessel occlusion and vessel injury.</p>
--	--

Details on Affected Device, to assist in identification of the product involved:	<p>Product involved: This letter applies to:</p> <table border="1" style="margin-left: auto; margin-right: auto;"><thead><tr><th>Item/Product Number</th><th>Lot Number</th></tr></thead><tbody><tr><td>SE18030100</td><td>L92505</td></tr><tr><td>SE18030100</td><td>L92835</td></tr><tr><td>SE18030100</td><td>L93029</td></tr><tr><td>SE18030100</td><td>L93495</td></tr></tbody></table> <p>Intended for Use: The SELUTION SLR™ PTA DEB is intended for use as a Percutaneous Transluminal Angioplasty (PTA) balloon catheter to dilate de nova or restenotic vascular lesions, for the purpose of improving limb perfusion and decreasing the incidence of restenosis.</p>	Item/Product Number	Lot Number	SE18030100	L92505	SE18030100	L92835	SE18030100	L93029	SE18030100	L93495
Item/Product Number	Lot Number										
SE18030100	L92505										
SE18030100	L92835										
SE18030100	L93029										
SE18030100	L93495										



Description of the problem:	<p><u>What is the issue?</u> Cordis has identified that there is potential for deflation difficulty or “<i>slow deflation</i>” of 4 lots of SELUTION SLR™ 018 PTA 3.0mm X 100mm balloon catheters.</p> <p><u>Why are we removing this product?</u> The potential impact of deflation difficulty or “<i>slow deflation</i>” includes increased procedure time, additional intervention, vessel occlusion and vessel injury.</p> <p><u>Is there any concern with the product already used successfully in procedures?</u> There is no concern with product that has been successfully used.</p> <p><u>What other actions is Cordis taking?</u> Cordis has identified the root cause and will take appropriate corrective actions. Only these devices from the four specific lots are impacted by the issue.</p>
Available Assistance:	If you have any questions regarding this field safety notice, please contact your local sales representative or local sales office, or Cordis at: GMB-CordisFieldAction@cordis.com
Additional Information:	<p><u>Regulatory Notification</u> The applicable regulatory agencies and notified body are being notified that Cordis is voluntarily taking this action.</p>

We know that you place high trust in our products, and we appreciate your cooperation in this matter. Cordis is committed to maintaining your confidence in the safety and quality of the products that Cordis supplies.

Respectfully yours,

Vice President, Quality & Design Assurance
Cordis
cc: Materials Director; Field Action Contact or Risk Manager